



When attempting to measure hardness manually, inconsistent pressures and/or speeds are used while pressing the indenter and presser foot. This causes a problem of deviation in measured results. To solve this problem, CL-150 Series constant pressure/speed loaders have been developed. Applicable standards refer to this type of loader as a 'Durometer Holder.'

Model	Applicable Hardness Tester	Pressing Load
CL-150L	Models JA, A, B, C, etc.	1000gf 1250gf
CL-150H	Models JC, D, etc.	5000gf
CL-150M	Above listed Models	1000gf 1250gf 5000gf